

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.59	
Mean	173.61	Mean	562.59
SD	3.87	SD	14.57
CV	2.23	CV	2.59
Day-Run	Value	Day-Run	Value
1 - 1	168.1	1 - 1	580.2
2 - 1	172	2 - 1	564
3 - 1	170	3 - 1	581.7
6 - 1	174.9	6 - 1	545.6
8 - 1	167	8 - 1	586
9 - 1	176	9 - 1	583
10 - 1	178.1	10 - 1	561
11 - 1	175.3	11 - 1	574
17 - 1	173.1	17 - 1	542.2
20 - 1	180	20 - 1	548
21 - 1	168.3	21 - 1	561.7
22 - 1	174.1	22 - 1	558.7
23 - 1	175	23 - 1	570
24 - 1	167.9	24 - 1	554.6
25 - 1	171.1	25 - 1	540
27 - 1	178	27 - 1	544
28 - 1	173	28 - 1	570
29 - 1	177.7	29 - 1	553.8
30 - 1	177.5	30 - 1	576.2
31 - 1	175	31 - 1	557

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	174.54	SD	3	Exp. Date	1403/1/17	Lot. No	035

