

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.8	
Mean	29.07	Mean	140.05
SD	1.02	SD	2.52
CV	3.51	CV	1.8
Day-Run	Value	Day-Run	Value
1 - 1	29.3	1 - 1	137.7
3 - 1	27.7	2 - 1	142
3 - 2	27.7	3 - 1	136
6 - 1	29.9	6 - 1	145
8 - 1	30	8 - 1	143
9 - 1	30.4	9 - 1	141
10 - 1	29.1	10 - 1	141
11 - 1	31.3	11 - 1	138
17 - 1	28	17 - 1	142.7
20 - 1	29.4	20 - 1	139
21 - 1	29	21 - 1	140.7
22 - 1	28.5	22 - 1	139.7
23 - 1	29.1	23 - 1	140
24 - 1	27.8	24 - 1	135.3
25 - 1	28.4	25 - 1	136.6
27 - 1	29.7	27 - 1	142
28 - 1	27.5	28 - 1	141.1
29 - 1	30	29 - 1	138.8
30 - 1	28.9	30 - 1	139
31 - 1	29.6	31 - 1	142.3

Westgard	تفسیر	SGPT ( ALT )	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	28.96	SD	0.99	Exp. Date	1403/1/17	Lot. No	035

