

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 2.11	
Mean	38.04	Mean	140.9
SD	1	SD	2.97
CV	2.62	CV	2.11
Day-Run	Value	Day-Run	Value
2 - 1	37.8	1 - 1	140.5
2 - 2	37.8	2 - 1	137
3 - 1	38.8	3 - 1	143
6 - 1	38	6 - 1	138
8 - 1	36.1	8 - 1	141.7
9 - 1	36.9	9 - 1	136.8
10 - 1	38.9	10 - 1	140
11 - 1	37.9	11 - 1	139.3
17 - 1	38.4	17 - 1	137
20 - 1	37.2	20 - 1	143
21 - 1	37.7	21 - 1	144.2
22 - 1	39	22 - 1	143.1
23 - 1	39	23 - 1	137
24 - 1	37	24 - 1	142.5
25 - 1	36.1	25 - 1	145
27 - 1	38.4	27 - 1	137
28 - 1	38.9	28 - 1	143
29 - 1	38.9	29 - 1	142.6
30 - 1	38.2	30 - 1	146
31 - 1	39.8	31 - 1	141.2

Westgard	تفسیر	SGOT ( AST )	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	37.85	SD	1.14	Exp. Date	1403/1/17	Lot. No	035

