

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	4.26
SD	0.08
CV	1.93
Day-Run	Value
1 - 1	4.3
2 - 1	4.2
3 - 1	4.2
6 - 1	4.2
8 - 1	4.2
9 - 1	4.1
10 - 1	4.2
11 - 1	4.3
17 - 1	4.1
20 - 1	4.3
21 - 1	4.3
22 - 1	4.3
23 - 1	4.3
24 - 1	4.3
25 - 1	4.3
27 - 1	4.4
28 - 1	4.3
29 - 1	4.3
30 - 1	4.4
31 - 1	4.2

High - CV : 1.68	
Mean	4.84
SD	0.08
CV	1.68
Day-Run	Value
1 - 1	5
2 - 1	4.9
3 - 1	4.8
6 - 1	4.8
8 - 1	4.9
9 - 1	4.9
10 - 1	4.8
11 - 1	4.7
17 - 1	4.7
20 - 1	4.8
21 - 1	4.8
22 - 1	4.8
23 - 1	4.9
24 - 1	4.8
25 - 1	4.8
27 - 1	4.8
28 - 1	4.8
29 - 1	5
30 - 1	4.9
31 - 1	4.8

Westgard	تفسیر	Albomin	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	4.25	SD	0.07	Exp. Date	1403/1/17	Lot. No	035

