

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 2.23	
Mean	4.22	Mean	4.98
SD	0.07	SD	0.11
CV	1.62	CV	2.23
Day-Run	Value	Day-Run	Value
1 - 1	4.1	1 - 1	4.8
2 - 1	4.2	2 - 1	4.9
7 - 1	4.3	8 - 1	4.9
8 - 1	4.2	9 - 1	4.9
9 - 1	4.1	11 - 1	5
11 - 1	4.2	12 - 1	4.9
12 - 1	4.2	14 - 1	4.8
14 - 1	4.3	15 - 1	5
15 - 1	4.2	16 - 1	4.9
16 - 1	4.2	17 - 1	5
17 - 1	4.2	18 - 1	5
18 - 1	4.2	19 - 1	4.9
19 - 1	4.3	21 - 1	5
21 - 1	4.3	22 - 1	5.1
22 - 1	4.3	23 - 1	5.2
23 - 1	4.2	24 - 1	5
24 - 1	4.3	25 - 1	5.1
25 - 1	4.2	26 - 1	5.1
26 - 1	4.2	28 - 1	5.1
28 - 1	4.3	29 - 1	5
29 - 1	4.3	29 - 2	4.8
30 - 1	4.1	30 - 1	5.1

Westgard	تفسیر	Albomin	آزمایش	BS-800 اتوآنالایزر	دستگاه	5	ماه	1402	سال
Normal Control	biorex	Mean	4.28	SD	0.11	Exp. Date	1403/1/17	Lot. No	035

