

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	78.5	
2 - 1	78.6	
3 - 1	82.7	
6 - 1	76.5	
8 - 1	81.7	
9 - 1	79	
10 - 1	76.8	
11 - 1	81.9	
17 - 1	78.6	
20 - 1	80.4	
21 - 1	76.5	
22 - 1	76.9	
23 - 1	82.1	
24 - 1	77.5	
25 - 1	81.4	
27 - 1	77	
28 - 1	82	
29 - 1	77.5	
30 - 1	81.3	
31 - 1	78	

Westgard	تفسیر	Amylase	آزمایش	BS-800 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	79.04	SD	2.24	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	79.25	SD	2.2	CV	2.77			

