

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	18.4	
3 - 1	17.9	
4 - 1	17.3	
5 - 1	18.3	
6 - 1	17.9	
7 - 1	18.89	
10 - 1	18.6	
11 - 1	18.9	
12 - 1	17.5	
13 - 1	16.9	
14 - 1	19.2	
15 - 1	17.8	
17 - 1	18.4	
17 - 2	17.2	
19 - 1	17	
20 - 1	19.1	
21 - 1	17	
22 - 1	18.4	
24 - 1	18.4	
25 - 1	17.5	
27 - 1	18.1	
27 - 2	17.9	
28 - 1	18.7	
29 - 1	17.3	
31 - 1	18.4	

Westgard	تفسیر	C4	آزمایش	BS-480 اتوآنالایزر	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	18.02	SD	0.61	Exp. Date	1402/5/11	Lot. No	641-01-030/2
Next Month	Mean	18.04	SD	0.69	CV	3.81			

