

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	18.5	
2 - 1	17	
3 - 1	18.2	
4 - 1	18.2	
6 - 1	17	
7 - 1	17.5	
8 - 1	18.1	
9 - 1	18.4	
10 - 1	18.6	
11 - 1	17.7	
13 - 1	18.2	
16 - 1	17.4	
17 - 1	17.8	
18 - 1	18.5	
20 - 1	17.2	
21 - 1	18.6	
22 - 1	16.9	
23 - 1	18.2	
24 - 1	18.5	
25 - 1	18.8	
27 - 1	17.3	
28 - 1	18	
29 - 1	18.6	
30 - 1	18.5	
31 - 1	18.9	

Westgard	تفسیر	C4	آزمایش	BS-480 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	17.93	SD	0.67	Exp. Date	1402/1/9	Lot. No	641-01-030/2
Next Month	Mean	18.02	SD	0.61	CV	3.38			

