

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	53.45	
2 - 1	54.5	
3 - 1	55.8	
4 - 1	55.24	
7 - 1	51.85	
8 - 1	55.4	
9 - 1	55.9	
10 - 1	52.5	
11 - 1	53.9	
12 - 1	55.4	
14 - 1	53	
15 - 1	49.5	
16 - 1	51.6	
17 - 1	53.9	
18 - 1	55.8	
19 - 1	51.5	
21 - 1	53.9	
22 - 1	45.9	
23 - 1	57.4	
24 - 1	46.5	
25 - 1	53.4	
26 - 1	54.7	
28 - 1	52.4	
29 - 1	51.9	
30 - 1	54.6	
31 - 1	55.8	

Westgard	تفسیر	CA 15-3	آزمایش	2000 ایمولایت XPI-	دستگاه	5	ماه	1402	سال
Control	Seronorm	Mean	53.35	SD	5.59	Exp. Date	1402/4/25	Lot. No	2011909
Next Month	Mean	53.3	SD	2.74	CV	5.15			

