

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.47	
Mean	127.5	Mean	215.26
SD	2.06	SD	3.16
CV	1.62	CV	1.47
Day-Run	Value	Day-Run	Value
1 - 1	126	1 - 1	219
2 - 1	128	2 - 1	217
7 - 1	128	7 - 1	216
8 - 1	125	8 - 1	215
9 - 1	130	9 - 1	210.8
11 - 1	126	11 - 1	215
12 - 1	131	12 - 1	216
14 - 1	128	14 - 1	219
15 - 1	131	15 - 1	218
16 - 1	125	16 - 1	215
17 - 1	125	17 - 1	217
18 - 1	127	18 - 1	210
19 - 1	127	19 - 1	212
21 - 1	126	21 - 1	218
22 - 1	129	22 - 1	218
23 - 1	126	23 - 1	221
24 - 1	126	24 - 1	216
25 - 1	127	25 - 1	210
26 - 1	125	26 - 1	215
28 - 1	130	28 - 1	210
29 - 1	131	29 - 1	215
30 - 1	128	30 - 1	213

Westgard	تفسیر	CPK	آزمایش	BS-800 اتوآنالایزر	دستگاه	5	ماه	1402	سال
Normal Control	biorex	Mean	126.68	SD	1.92	Exp. Date	1403/1/17	Lot. No	035

