

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	49.1	
3 - 1	51	
4 - 1	49.7	
5 - 1	49.2	
6 - 1	51.3	
7 - 1	49.9	
10 - 1	52.9	
10 - 2	54	
12 - 1	50.7	
13 - 1	51.3	
14 - 1	51.3	
15 - 1	48.5	
17 - 1	50.3	
18 - 1	48.5	
19 - 1	52	
20 - 1	48.2	
21 - 1	48.4	
22 - 1	50.8	
24 - 1	47.6	
25 - 1	49.5	
26 - 1	50.8	
28 - 1	52.9	
28 - 2	49.4	
31 - 1	50.9	

Westgard	تفسیر	CRP	آزمایش	BS-480 اتوآنالایزر	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	50.86	SD	1.68	Exp. Date	1402/5/1	Lot. No	326-2-034/3
Next Month	Mean	50.34	SD	1.63	CV	3.23			

