

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
10 - 1	25.5	
11 - 1	22.9	
13 - 1	24.3	
16 - 1	26.1	
18 - 1	26.5	
20 - 1	24	
21 - 1	25.3	
22 - 1	25.1	
23 - 1	26.2	
24 - 1	25	
25 - 1	26	
27 - 1	24.1	
28 - 1	24.5	
29 - 1	24.8	
30 - 1	26.4	
31 - 1	23	

Westgard	تفسیر	Ceruloplasmin	آزمایش	BS-480 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	24.7	SD	1.14	Exp. Date	1402/1/9	Lot. No	641-01-030/2
Next Month	Mean	24.98	SD	1.13	CV	4.53			

