

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	176.5
SD	2.41
CV	1.36
Day-Run	Value
1 - 1	178
2 - 1	176
7 - 1	180
8 - 1	180
9 - 2	175
11 - 1	180
12 - 1	175
14 - 1	177
15 - 1	179
16 - 1	175
17 - 1	172
18 - 1	175
19 - 1	180
21 - 1	177
22 - 1	174
23 - 1	175
24 - 1	175
25 - 1	179
26 - 1	174
28 - 1	174
29 - 1	175
30 - 1	178

High - CV : 0.73	
Mean	263.95
SD	1.94
CV	0.73
Day-Run	Value
1 - 1	265
2 - 1	266
7 - 1	265
8 - 1	265
9 - 1	262
11 - 1	262
12 - 1	262
14 - 1	264
15 - 1	265
16 - 1	260
17 - 1	265
18 - 1	262
19 - 1	265
21 - 1	266
22 - 1	263
23 - 1	262
24 - 1	263
25 - 1	266
26 - 1	267
28 - 1	262
29 - 1	263
30 - 1	267

Westgard	تفسیر	Cholesterol	آزمایش	اتوآنالایزر BS-800	دستگاه	5	ماه	1402	سال
Normal Control	biorex	Mean	176	SD	2.74	Exp. Date	1403/1/17	Lot. No	035

