

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.33	
Mean	2.55	Mean	3.63
SD	0.05	SD	0.08
CV	2.07	CV	2.33
Day-Run	Value	Day-Run	Value
1 - 1	2.58	1 - 1	3.62
3 - 1	2.6	3 - 1	3.67
4 - 1	2.47	4 - 1	3.56
5 - 1	2.48	5 - 1	3.55
6 - 1	2.48	6 - 1	3.6
10 - 1	2.55	10 - 1	3.65
11 - 1	2.44	11 - 1	3.58
13 - 1	2.48	13 - 1	3.55
19 - 1	2.6	19 - 1	3.75
20 - 1	2.56	20 - 1	3.66
21 - 1	2.59	21 - 1	3.74
24 - 1	2.6	24 - 1	3.41
25 - 1	2.53	25 - 1	3.7
26 - 1	2.57	26 - 1	3.68
27 - 1	2.6	27 - 1	3.75
28 - 1	2.56	28 - 1	3.64
29 - 1	2.54	29 - 1	3.63
31 - 1	2.58	31 - 1	3.66

Westgard	تفسیر	Creatinine	آزمایش	BS-800 اتوآنالایزر	دستگاه	4	ماه	1402	سال
Normal Control	biorex	Mean	2.52	SD	0.06	Exp. Date	1403/1/17	Lot. No	035

