

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 5.02	
Mean	0.46	Mean	2.2
SD	0.03	SD	0.11
CV	6.4	CV	5.02
Day-Run	Value	Day-Run	Value
1 - 1	0.43	1 - 1	2.24
2 - 1	0.48	2 - 1	2.28
7 - 1	0.46	7 - 1	2.15
8 - 1	0.49	8 - 1	2.38
9 - 1	0.48	9 - 1	2
11 - 1	0.45	11 - 1	2.3
12 - 1	0.5	12 - 1	2.13
14 - 1	0.42	14 - 1	2.3
15 - 1	0.42	15 - 1	2.05
16 - 1	0.44	16 - 1	2.1
17 - 1	0.47	17 - 1	2.25
18 - 1	0.46	18 - 1	2.07
19 - 1	0.5	19 - 1	2.3
21 - 1	0.5	21 - 1	2.05
22 - 1	0.46	22 - 1	2.1
23 - 1	0.43	23 - 1	2.23
24 - 1	0.47	24 - 1	2.18
25 - 1	0.52	25 - 1	2.21
26 - 1	0.46	26 - 1	2.15
28 - 1	0.47	28 - 1	2.3
29 - 1	0.42	29 - 1	2.2
30 - 1	0.43	30 - 1	2.4

Westgard	تفسیر	D.Bili	آزمایش	BS-800 اتوآنالایزر	دستگاه	5	ماه	1402	سال
Normal Control	biorex	Mean	0.45	SD	0.03	Exp. Date	1403/1/17	Lot. No	035

