

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.1	
Mean	75.65	Mean	228.7
SD	1.09	SD	4.81
CV	1.44	CV	2.1
Day-Run	Value	Day-Run	Value
1 - 1	76	1 - 1	238
2 - 1	76	2 - 1	234
3 - 1	76	3 - 1	238
6 - 1	75	6 - 1	230
8 - 1	76	8 - 1	229
9 - 1	76	9 - 1	229
10 - 1	75	10 - 1	232
11 - 1	76	11 - 1	223
17 - 1	75	17 - 1	225
20 - 1	77	20 - 1	223
21 - 1	76	21 - 1	233
22 - 1	75	22 - 1	225
23 - 1	76	23 - 1	226
24 - 1	74	24 - 1	231
25 - 1	75	25 - 1	228
27 - 1	79	27 - 1	231
28 - 1	74	28 - 1	220
29 - 1	75	29 - 1	225
30 - 1	75	30 - 1	228
31 - 1	76	31 - 1	226

Westgard	تفسیر	FBS	آزمایش	BS-800 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	75	SD	2.34	Exp. Date	1403/1/17	Lot. No	035

