

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	7.42	
2 - 1	7.56	
3 - 1	7.62	
4 - 1	7.45	
7 - 1	7.35	
8 - 1	7.63	
9 - 1	7.35	
10 - 1	7.45	
11 - 1	7.5	
14 - 1	7.37	Cusum 3S
15 - 1	7.35	
16 - 1	7.45	
17 - 1	7.62	
18 - 1	7.46	1:2S
19 - 1	7.42	
21 - 1	7.35	
22 - 1	7.65	
23 - 1	7.59	
24 - 1	7.45	
25 - 1	7.46	
26 - 1	7.65	
28 - 1	7.35	
29 - 1	7.35	
30 - 1	7.42	
31 - 1	7.68	

Westgard	تفسیر	Free T3	آزمایش	SIEMENS Centaur CP	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	7.5	SD	0.1	Exp. Date	1402/10/30	Lot. No	36-02-026
Next Month	Mean	7.48	SD	0.11	CV	1.51			

