

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	67.5	1:2S
2 - 1	68.3	
7 - 1	70.9	
8 - 1	69.2	
9 - 1	67	
11 - 1	69.8	
12 - 1	69	
14 - 1	71.2	
15 - 1	68.6	
16 - 1	67.4	
17 - 1	70	
18 - 1	69.7	
19 - 1	68	
21 - 1	66.9	
23 - 1	70	
24 - 1	69.3	
25 - 1	66.4	
26 - 1	66.8	
28 - 1	68	
29 - 1	69	
30 - 1	69	

Westgard	تفسیر	Gamma GT	آزمایش	اتوآنالایزر BS-800	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	69.07	SD	1.54	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	68.67	SD	1.36	CV	1.98			

