

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	12.7	
2 - 1	12.7	
3 - 1	13	
4 - 1	12.8	
6 - 1	12.5	
7 - 1	12.5	
8 - 1	12.5	
9 - 1	12.6	
10 - 1	13	
11 - 1	12.8	
16 - 1	12.5	
17 - 1	12.8	
18 - 1	12.5	
20 - 1	12.9	
21 - 1	12.9	
23 - 1	13	
24 - 1	13	
25 - 1	13	
27 - 1	13.1	4:1S

Westgard	تفسیر	HB	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	3	ماه	1402	سال
Control	Labex	Mean	12.75	SD	0.23	Exp. Date	1402/2/9	Lot. No	
Next Month	Mean	12.77	SD	0.21	CV	1.61			

