

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.42	
Mean	12.14	Mean	17.52
SD	0.15	SD	0.25
CV	1.27	CV	1.42
Day-Run	Value	Day-Run	Value
2 - 1	12	2 - 1	17.4
2 - 2	11.9	2 - 2	17.5
3 - 1	12	3 - 1	17.4
4 - 1	12.1	4 - 1	17.6
8 - 1	12.4	8 - 1	17.4
9 - 1	12.1	9 - 1	17.5
11 - 1	12.1	11 - 1	17
12 - 1	12	12 - 1	17.5
15 - 1	12	15 - 1	17.6
17 - 1	12.3	17 - 1	17.7
18 - 1	12.1	18 - 1	17.8
21 - 1	12.1	21 - 1	17.3
22 - 1	12.2	22 - 1	17.9
23 - 1	12.2	23 - 1	17.5
24 - 1	12.4	24 - 1	17.9
25 - 1	12.3	25 - 1	17.7
28 - 1	12.1	28 - 1	17
29 - 1	12	29 - 1	17.5
31 - 1	12.4	31 - 1	17.7

Westgard	تفسیر	HB	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	5	ماه	1402	سال
Normal Control	Labex	Mean	12.09	SD	0.21	Exp. Date	1402/6/30	Lot. No	02-2305n

