

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	41.6	
2 - 1	42.1	
3 - 1	42	
4 - 1	42.5	1:3S
6 - 1	42	
7 - 1	42	
8 - 1	41.6	
9 - 1	41.5	
10 - 1	42	
13 - 1	42	
16 - 1	41.5	
16 - 2	41	
17 - 1	41.2	
18 - 1	42	
19 - 1	41.5	
20 - 1	41.8	
21 - 1	42.1	
23 - 1	42	
24 - 1	41.3	
25 - 1	42.4	
27 - 1	42	

Westgard	تفسیر	HCT	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	3	ماه	1402	سال
Control	Labex	Mean	41.7	SD	0.59	Exp. Date	1402/2/9	Lot. No	
Next Month	Mean	41.81	SD	0.38	CV	0.92			

