

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	517	
2 - 1	538.2	
3 - 1	539.9	
4 - 1	550.9	
6 - 1	552.1	
7 - 1	545.5	
8 - 1	543.4	
9 - 1	532.2	
10 - 1	540.5	
11 - 1	560.7	
13 - 1	554.2	
16 - 1	569.5	
17 - 1	520.5	
18 - 1	518.5	
20 - 1	530.7	
21 - 1	554	
22 - 1	556.9	
23 - 1	521.6	
24 - 1	539.8	
25 - 1	536.7	
27 - 1	533.7	
28 - 1	520.7	
29 - 1	528.5	
30 - 1	532.2	
31 - 1	502.1	

Westgard	تفسیر	Ig G	آزمایش	BS-480 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	534.82	SD	18.05	Exp. Date	1402/1/30	Lot. No	641-01-030/2
Next Month	Mean	537.6	SD	15.89	CV	2.96			

