

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
10 - 1	46.4	
11 - 1	41.3	
13 - 1	42	
16 - 1	48.3	
17 - 1	44.4	
18 - 1	47.3	
20 - 1	42	
21 - 1	45.2	
22 - 1	46.3	
23 - 1	47.7	
24 - 1	47.7	
25 - 1	47.5	
27 - 1	41.7	
28 - 1	44.1	
29 - 1	43.2	
30 - 1	43.6	
31 - 1	42.8	

Westgard	تفسیر	Ig M	آزمایش	BS-480 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	45.2	SD	2.41	Exp. Date	1402/1/30	Lot. No	641-01-030/2
Next Month	Mean	44.79	SD	2.42	CV	5.4			

