

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	66.2	
2 - 1	62.7	
3 - 1	64	
4 - 1	57.5	
7 - 1	56.9	
8 - 1	70.15	Cusum 3S
9 - 1	65.2	
10 - 1	70.5	
11 - 1	57.31	
14 - 1	55.25	
15 - 1	58	
16 - 1	70.5	
17 - 1	68.2	
18 - 1	70.52	
19 - 1	56.2	
21 - 1	58	
22 - 1	65.9	
23 - 1	70.51	
24 - 1	58	
25 - 1	58.5	
26 - 1	53.7	
28 - 1	50.8	
29 - 1	66.9	
30 - 1	68.3	
31 - 1	53.9	

Westgard	تفسیر	Insulin (Base)	آزمایش	SIEMENS Centaur CP	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	58.8	SD	6	Exp. Date	1402/10/30	Lot. No	36-02-028
Next Month	Mean	62.15	SD	6.37	CV	10.24			

