

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.11	
3 - 1	4.22	
4 - 1	4.18	
5 - 1	4.1	
6 - 1	4.15	
7 - 1	4.19	
10 - 1	4.15	
11 - 1	4.21	
12 - 1	4.11	
13 - 1	4.16	
14 - 1	4.2	
15 - 1	4.15	
17 - 1	4.1	
18 - 1	4.17	
19 - 1	4.12	
20 - 1	4.23	
21 - 1	4.15	
22 - 1	4.13	
24 - 1	4.1	
25 - 1	4.19	
27 - 1	4.15	
27 - 2	4.11	
28 - 1	4.18	
29 - 1	4.16	
31 - 1	4.12	

Westgard	تفسیر	Potassium ( K )	آزمایش	AUDICOM	دستگاه	4	ماه	1402	سال
Control	Ringer	Mean	4.15	SD	0.04	Exp. Date	1402/5/11	Lot. No	761LOT17
Next Month	Mean	4.15	SD	0.04	CV	0.95			

