

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 2.59	
Mean	313.05	Mean	441.05
SD	9.45	SD	11.42
CV	3.02	CV	2.59
Day-Run	Value	Day-Run	Value
1 - 1	303	1 - 1	442.4
2 - 1	314	2 - 1	453.4
7 - 1	305	7 - 1	442
8 - 1	301	8 - 1	434
9 - 1	315	9 - 1	453
11 - 1	316	12 - 1	423
12 - 1	301	14 - 1	431
15 - 1	319	15 - 1	461.3
16 - 1	318	16 - 1	430
17 - 1	306	17 - 1	460
18 - 1	320	18 - 1	430
19 - 1	300	19 - 1	430
21 - 1	303	21 - 1	458
22 - 1	327	22 - 1	442
23 - 1	313	23 - 1	430
24 - 1	305	24 - 1	433
25 - 1	319	25 - 1	448
26 - 1	312	26 - 1	449
28 - 1	333	28 - 1	435
29 - 1	326	29 - 1	444
30 - 1	318	30 - 1	433

Westgard	تفسیر	LDH	آزمایش	اتوآنالایزر BS-800	دستگاه	5	ماه	1402	سال
Normal Control	biorex	Mean	311.49	SD	7.68	Exp. Date	1403/1/17	Lot. No	035

