

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	134	
2 - 1	137	
7 - 1	135	
8 - 1	138	
9 - 1	140	
11 - 1	138	
12 - 1	133	
14 - 1	137.2	
15 - 1	136	
16 - 1	138	
17 - 1	140.1	
18 - 1	137.4	
19 - 1	140	
21 - 1	132	
22 - 1	137.6	
23 - 1	140	
24 - 1	140	
25 - 1	137.6	
26 - 1	133.7	
28 - 1	140	
29 - 1	134.5	
30 - 1	139	

Westgard	تفسیر	LDL	آزمایش	BS-800 اتوآنالایزر	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	136.93	SD	2.52	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	137.19	SD	2.53	CV	1.84			

