

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 0.71	
Mean	28.07	Mean	31.71
SD	0.36	SD	0.22
CV	1.3	CV	0.71
Day-Run	Value	Day-Run	Value
2 - 1	27.7	2 - 1	31.5
2 - 2	27.9	2 - 2	31.8
3 - 1	28	3 - 1	32
4 - 1	28.2	4 - 1	31.5
8 - 1	28.3	8 - 1	31.6
9 - 1	27.8	9 - 1	31.9
11 - 1	28.2	11 - 1	31.5
12 - 1	27.5	12 - 1	32
15 - 1	27.8	15 - 1	31.6
17 - 1	28.2	17 - 1	32
18 - 1	28.3	18 - 1	32
21 - 1	27.5	21 - 1	31.8
22 - 1	28.5	22 - 1	31.5
23 - 1	27.9	23 - 1	32
24 - 1	29	24 - 1	31.6
25 - 1	28.5	25 - 1	31.6
28 - 1	28	28 - 1	31.7
29 - 1	28	29 - 1	31.3
31 - 1	28.1	31 - 1	31.5

Westgard	تفسیر	MCH	آزمایش	NIHON KOHDEN سل کانتیر	دستگاه	5	ماه	1402	سال
Normal Control	Labex	Mean	28.16	SD	0.44	Exp. Date	1402/6/30	Lot. No	02-2305n

