

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	30.58
SD	0.59
CV	1.93
Day-Run	Value
1 - 1	30.4
3 - 1	32.3
4 - 1	30.7
5 - 1	30.4
6 - 1	30.7
7 - 1	31
10 - 1	30.8
11 - 1	30.2
12 - 1	30.5
13 - 1	30.3
14 - 1	30.8
15 - 1	30.9
17 - 1	30.8
18 - 1	30.9
19 - 1	30.6
20 - 1	31
18 - 2	30.5
20 - 2	31.3
21 - 1	30.2
22 - 1	30
24 - 1	30.8
25 - 1	29.1
26 - 1	29.7
28 - 1	30
29 - 1	30.5

High - CV : 1.11	
Mean	31.68
SD	0.35
CV	1.11
Day-Run	Value
1 - 1	31.6
3 - 1	31.9
4 - 1	31.7
5 - 1	31.9
6 - 1	31.6
7 - 1	32.2
10 - 1	31.6
11 - 1	31.4
12 - 1	31.2
13 - 1	31.3
14 - 1	32.1
15 - 1	31.5
17 - 1	32
18 - 1	31
19 - 1	31.7
20 - 1	31.4
18 - 2	31.6
21 - 1	32.4
22 - 1	31.3
24 - 1	32.2
25 - 1	31.8
26 - 1	31.9
27 - 1	31.3
28 - 1	31.5
29 - 1	32

Westgard	تفسیر	MCHC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	4	ماه	1402	سال
Normal Control		Mean	30.76	SD	0.58	Exp. Date	1402/4/13	Lot. No	

