

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 0.65	
Mean	91.65	Mean	99.13
SD	0.64	SD	0.64
CV	0.7	CV	0.65
Day-Run	Value	Day-Run	Value
2 - 1	90	2 - 1	99
2 - 2	91.5	2 - 2	99
3 - 1	92	3 - 1	99.4
4 - 1	92	4 - 1	99
8 - 1	90.8	8 - 1	98.2
9 - 1	92.2	9 - 1	99.5
11 - 1	91	11 - 1	99.4
12 - 1	92.3	12 - 1	99.5
15 - 1	91.4	15 - 1	98.2
17 - 1	92.2	17 - 1	99.3
18 - 1	91.6	18 - 1	100
21 - 1	91.7	21 - 1	99.4
22 - 1	91.4	22 - 1	99.6
23 - 1	91.6	23 - 1	99
24 - 1	91.8	24 - 1	100
25 - 1	92.8	25 - 1	100
28 - 1	91	28 - 1	98
29 - 1	92	29 - 1	99
31 - 1	92	31 - 1	98

Westgard	تفسیر	MCV	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	5	ماه	1402	سال
Normal Control	Labex	Mean	91.83	SD	0.96	Exp. Date	1402/6/30	Lot. No	02-2305n

