

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	2.24
SD	0.13
CV	5.62
Day-Run	Value
1 - 1	2.3
2 - 1	2.2
3 - 1	2.2
6 - 1	2.1
8 - 1	2.3
9 - 1	2.3
10 - 1	2.2
11 - 1	2.1
17 - 1	2.2
20 - 1	2.1
21 - 1	2.4
22 - 1	2.4
23 - 1	2.2
24 - 1	2.5
25 - 1	2.3
27 - 1	2
29 - 1	2.1
30 - 1	2.3
31 - 1	2.3

High - CV : 4.92	
Mean	3.82
SD	0.19
CV	4.92
Day-Run	Value
1 - 1	4.1
2 - 1	4
3 - 1	3.7
6 - 1	3.7
8 - 1	4
9 - 1	3.9
10 - 1	3.7
11 - 1	3.5
17 - 1	4
20 - 1	3.6
21 - 1	3.9
22 - 1	4.1
23 - 1	3.9
24 - 1	3.7
25 - 1	3.8
27 - 1	3.6
28 - 1	4
29 - 1	3.8
30 - 1	3.9
31 - 1	3.5

Westgard	تفسیر	Magnesium(Mg)	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	2.2	SD	0.12	Exp. Date	1403/1/17	Lot. No	035

