

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	152	
3 - 1	152.8	
5 - 1	151.2	
6 - 1	152.8	
7 - 1	151.5	
10 - 1	153	
11 - 1	152.4	
12 - 1	153.2	
13 - 1	153	
14 - 1	152.8	
15 - 1	151.5	
17 - 1	153	
18 - 1	152	
19 - 1	151.4	
20 - 1	153.5	
21 - 1	152.2	
22 - 1	151.5	
24 - 1	152.5	
25 - 1	153	
26 - 1	152	
27 - 1	153.2	
29 - 1	153	
31 - 1	152	

Westgard	تفسیر	Sodium (Na)	آزمایش	AUDICOM	دستگاه	4	ماه	1402	سال
Control	Ringer	Mean	152.14	SD	0.8	Exp. Date	1402/5/11	Lot. No	761LOT17
Next Month	Mean	152.41	SD	0.69	CV	0.45			

