

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.19	
Mean	223.39	Mean	549.56
SD	3.76	SD	12.03
CV	1.68	CV	2.19
Day-Run	Value	Day-Run	Value
2 - 1	228	2 - 1	530
2 - 2	225	2 - 2	550
3 - 1	225	3 - 1	560
4 - 1	220	4 - 1	550
8 - 1	219	8 - 1	552
9 - 1	227	9 - 1	532
11 - 1	218	11 - 1	560
12 - 1	225	12 - 1	555
15 - 1	220	15 - 1	530
17 - 1	225	17 - 1	556
18 - 1	230	18 - 1	545
22 - 1	225	22 - 1	532
23 - 1	218	23 - 1	546
24 - 1	225	24 - 1	555
25 - 1	219	25 - 1	560
28 - 1	227	28 - 1	570
29 - 1	220	29 - 1	563
31 - 1	225	31 - 1	546

Westgard	تفسیر	PLT	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	5	ماه	1402	سال
Normal Control	Labex	Mean	222.39	SD	4.65	Exp. Date	1402/6/30	Lot. No	02-2305n

