

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	10.25	
2 - 1	10.52	
3 - 1	9.56	
4 - 1	10	
7 - 1	9.57	
8 - 1	10.6	
9 - 1	11	
10 - 1	10.82	
11 - 1	9.48	
14 - 1	9.25	
15 - 1	9.5	
16 - 1	10.25	
17 - 1	9.52	Cusum 3S
18 - 1	9.26	
19 - 1	10.56	
21 - 1	10.25	
22 - 1	9.25	
23 - 1	9.72	
24 - 1	9.35	
25 - 1	10.9	
26 - 1	11	
28 - 1	9.73	
29 - 1	9.35	
30 - 1	9.78	
31 - 1	9.66	

Westgard	تفسیر	Progesterone	آزمایش	SIEMENS Centaur CP	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	10.17	SD	0.48	Exp. Date	1402/10/30	Lot. No	36-02-028
Next Month	Mean	9.97	SD	0.6	CV	5.99			

