

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	4.32
SD	0.06
CV	1.28
Day-Run	Value
1 - 1	4.34
3 - 1	4.34
4 - 1	4.44
5 - 1	4.24
6 - 1	4.33
7 - 1	4.3
10 - 1	4.3
11 - 1	4.28
12 - 1	4.21
13 - 1	4.28
14 - 1	4.3
15 - 1	4.35
16 - 1	4.39
17 - 1	4.41
18 - 1	4.36
20 - 1	4.32
21 - 1	4.24
22 - 1	4.32
24 - 1	4.26
25 - 1	4.29
26 - 1	4.34
27 - 1	4.28
28 - 1	4.28
29 - 1	4.37

High - CV : 2.33	
Mean	5.52
SD	0.13
CV	2.33
Day-Run	Value
1 - 1	5.51
3 - 1	5.5
4 - 1	5.52
5 - 1	5.47
6 - 1	5.47
7 - 1	5.54
10 - 1	5.49
11 - 1	5.48
12 - 1	5.51
13 - 1	5.48
14 - 1	5.92
15 - 1	5.9
16 - 1	5.44
17 - 1	5.44
18 - 1	5.49
21 - 1	5.42
22 - 1	5.58
24 - 1	5.45
25 - 1	5.53
26 - 1	5.42
27 - 1	5.51
28 - 1	5.53
29 - 1	5.43

Westgard	تفسیر	RBC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	4	ماه	1402	سال
Normal Control		Mean	4.31	SD	0.06	Exp. Date	1402/4/13	Lot. No	

