

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.36	1:2S
2 - 1	4.43	Cusum 3S
3 - 1	4.45	
4 - 1	4.4	
6 - 1	4.42	
7 - 1	4.43	
8 - 1	4.45	
9 - 1	4.42	
10 - 1	4.41	
11 - 1	4.41	
13 - 1	4.43	
16 - 1	4.45	
17 - 1	4.44	
18 - 1	4.4	
20 - 1	4.47	
21 - 1	4.44	
23 - 1	4.45	
24 - 1	4.39	
25 - 1	4.48	
27 - 1	4.48	
28 - 1	4.43	

Westgard	تفسیر	RBC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	3	ماه	1402	سال
Control	Labex	Mean	4.44	SD	0.03	Exp. Date	1402/2/9	Lot. No	
Next Month	Mean	4.43	SD	0.03	CV	0.67			

