

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.44	
1 - 2	2.64	
3 - 1	2.5	
4 - 1	2.55	
5 - 1	2.36	
6 - 1	2.16	
7 - 1	2.5	
10 - 1	2.66	1:3S
11 - 1	2.61	
12 - 1	2.53	
13 - 1	2.4	
14 - 1	2.4	
15 - 1	2.53	
17 - 1	2.57	
18 - 1	2.51	
19 - 1	2.3	
20 - 1	2.7	
21 - 1	2.31	
22 - 1	2.25	
24 - 1	2.4	
25 - 1	2.7	
26 - 1	2.65	
27 - 1	2.2	
28 - 1	2.2	
29 - 1	2.4	
31 - 1	2.55	
1 - 3	2.48	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	2.43	SD	0.15	Exp. Date	1402/5/1	Lot. No	36-02-028
Next Month	Mean	2.46	SD	0.16	CV	6.33			

