

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.52	
2 - 1	2.6	
3 - 1	2.4	
4 - 1	2.35	
6 - 1	2.29	
7 - 1	2.45	
8 - 1	2.75	
9 - 1	2.6	
10 - 1	2.65	
11 - 1	2.3	
13 - 1	2.49	
16 - 1	2.57	
17 - 1	2.5	
18 - 1	2.7	
20 - 1	2.32	
21 - 1	2.26	
22 - 1	2.65	
23 - 1	2.35	
24 - 1	2.7	
25 - 1	2.58	
27 - 1	2.55	
28 - 1	2.65	
29 - 1	2.3	
30 - 1	2.42	
31 - 1	2.5	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	2.51	SD	0.15	Exp. Date	1402/8/1	Lot. No	365-02-026
Next Month	Mean	2.5	SD	0.15	CV	5.91			

