

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.48	
2 - 1	2.53	
3 - 1	2.42	1:3S Cusum 3S
4 - 1	2.26	
7 - 1	2.37	
8 - 1	2.57	
9 - 1	2.26	
10 - 1	2.27	Cusum 3S
11 - 1	2.29	
14 - 1	2.51	
15 - 1	2.35	1:2S
16 - 1	2.33	
17 - 1	2.75	
18 - 1	2.58	
19 - 1	2.6	
21 - 1	2.45	
22 - 1	2.75	
23 - 1	2.66	
24 - 1	2.58	
25 - 1	2.54	
26 - 1	2.38	
28 - 1	2.61	
29 - 1	2.38	
30 - 1	2.7	
31 - 1	2.63	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	2.46	SD	0.16	Exp. Date	1402/10/30	Lot. No	36-02-028
Next Month	Mean	2.49	SD	0.15	CV	6.17			

