

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	10.3	
1 - 2	11	
3 - 1	10.1	
4 - 1	10.7	
5 - 1	11.2	
6 - 1	10.6	
7 - 1	10.1	
10 - 1	10.5	
11 - 1	11.1	
12 - 1	10.9	
13 - 1	10.8	
14 - 1	10.3	
15 - 1	11.22	
17 - 1	11.9	
18 - 1	11.52	
19 - 1	10.2	
20 - 1	9.75	
21 - 1	10.5	1:3S
22 - 1	11.52	
24 - 1	11.35	
25 - 1	9.7	
26 - 1	10.2	
27 - 1	10.1	
28 - 1	11.32	
29 - 1	12	
31 - 1	10.6	

Westgard	تفسیر	T4	آزمایش	SIEMENS Centaur CP	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	11.08	SD	0.71	Exp. Date	1402/5/1	Lot. No	36-02-028
Next Month	Mean	10.75	SD	0.63	CV	5.87			

