

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	11.42	
2 - 1	11.5	
3 - 1	11.68	
4 - 1	12.5	
6 - 1	11.97	
7 - 1	11.25	
8 - 1	12.6	
9 - 1	12.57	
10 - 1	11.6	
11 - 1	11.5	
13 - 1	12	
16 - 1	12.35	
17 - 1	12.5	
18 - 1	11.5	
20 - 1	11.8	
21 - 1	11.45	
22 - 1	12.35	
23 - 1	12.5	
24 - 1	12.45	
25 - 1	11.55	
27 - 1	11.52	
28 - 1	11.3	
29 - 1	12	
30 - 1	12.3	
31 - 1	11.7	

Westgard	تفسیر	T4	آزمایش	SIEMENS Centaur CP	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	11.89	SD	0.42	Exp. Date	1402/8/1	Lot. No	365-02-026
Next Month	Mean	11.91	SD	0.46	CV	3.85			

