

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.58	
Mean	172.1	Mean	223.8
SD	2.47	SD	3.53
CV	1.43	CV	1.58
Day-Run	Value	Day-Run	Value
1 - 1	173	1 - 1	230
2 - 1	170	2 - 1	230
3 - 1	177	3 - 1	226
6 - 1	171	6 - 1	227
8 - 1	173	8 - 1	227
9 - 1	171	9 - 1	227
10 - 1	174	10 - 1	225
11 - 1	176	11 - 1	223
17 - 1	169	17 - 1	222
20 - 1	171	20 - 1	221
21 - 1	175	21 - 1	225
22 - 1	170	22 - 1	217
23 - 1	170	23 - 1	222
24 - 1	170	24 - 1	225
25 - 1	175	25 - 1	220
27 - 1	173	27 - 1	222
28 - 1	168	28 - 1	218
29 - 1	171	29 - 1	224
30 - 1	171	30 - 1	221
31 - 1	174	31 - 1	224

Westgard	تفسیر	Triglycerides	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	173.47	SD	2.21	Exp. Date	1403/1/17	Lot. No	035

