

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	3.5	
1 - 2	3.8	
3 - 1	3.37	
4 - 1	3.52	
5 - 1	3.37	
6 - 1	3.61	
7 - 1	3.65	
10 - 1	3.57	
11 - 1	3.74	
12 - 1	3.71	
13 - 1	3.36	
14 - 1	3.37	
15 - 1	3.42	
17 - 1	3.69	
18 - 1	3.44	
19 - 1	3.39	
20 - 1	3.56	
21 - 1	3.67	
22 - 1	3.7	
24 - 1	3.39	
25 - 1	3.35	
26 - 1	3.65	
27 - 1	3.4	
28 - 1	3.37	
29 - 1	3.72	
31 - 1	3.39	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	3.6	SD	0.15	Exp. Date	1402/5/1	Lot. No	36-02-028
Next Month	Mean	3.53	SD	0.15	CV				

