

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.35	
2 - 1	2.4	
3 - 1	2.55	
4 - 1	2.45	
6 - 1	2.52	
7 - 1	2.45	
8 - 1	2.35	
9 - 1	2.4	
10 - 1	2.5	
11 - 1	2.45	
13 - 1	2.4	
16 - 1	2.35	
17 - 1	2.37	
18 - 1	2.5	
20 - 1	2.53	
21 - 1	2.45	
22 - 1	2.3	
23 - 1	2.45	
24 - 1	2.5	
25 - 1	2.44	
27 - 1	2.32	
28 - 1	2.4	
29 - 1	2.35	
30 - 1	2.45	
31 - 1	2.5	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	3	ماه	1402	سال
Control	biorex	Mean	2.42	SD	0.07	Exp. Date	1402/8/1	Lot. No	36-02-026
Next Month	Mean	2.43	SD	0.07	CV	2.87			

