

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 1 - 1 | 3.68 | |
| 2 - 1 | 3.55 | |
| 3 - 1 | 3.38 | |
| 4 - 1 | 3.41 | |
| 7 - 1 | 3.5 | |
| 8 - 1 | 3.65 | |
| 9 - 1 | 3.81 | |
| 10 - 1 | 3.79 | Cusum 3S |
| 11 - 1 | 3.35 | |
| 14 - 1 | 3.42 | |
| 15 - 1 | 3.82 | |
| 16 - 1 | 3.72 | |
| 17 - 1 | 3.73 | |
| 18 - 1 | 3.45 | |
| 19 - 1 | 3.25 | |
| 21 - 1 | 3.33 | |
| 22 - 1 | 3.66 | |
| 23 - 1 | 3.39 | |
| 24 - 1 | 3.48 | |
| 25 - 1 | 3.63 | |
| 26 - 1 | 3.79 | |
| 28 - 1 | 3.8 | |
| 29 - 1 | 3.47 | |
| 30 - 1 | 3.8 | |
| 31 - 1 | 3.76 | |

| | | | | | | | | | |
|------------|--------|------|--------|--------------------|--------|-----------|------------|---------|-----------|
| Westgard | تفسیر | TSH | آزمایش | SIEMENS Centaur CP | دستگاه | 5 | ماه | 1402 | سال |
| Control | biorex | Mean | 3.53 | SD | 0.15 | Exp. Date | 1402/10/30 | Lot. No | 36-02-028 |
| Next Month | Mean | 3.58 | SD | 0.18 | CV | 5.08 | | | |

