

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
24 - 1	380	
25 - 1	395	
26 - 1	389	
28 - 1	400	1:2S
29 - 1	448.3	
30 - 1	483.2	
31 - 1	471	
1 - 1	399.2	
2 - 1	492	
4 - 1	431	
5 - 1	389	
6 - 1	400	
7 - 1	390	
8 - 1	359	

Westgard	تفسیر	Testosterone	آزمایش	Maglumi	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	408.88	SD	42.08	Exp. Date	1403/5/19	Lot. No	365-02-028
Next Month	Mean	416.19	SD	41.6	CV	10			

