

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	144.6	
3 - 1	155.2	
4 - 1	147.9	
5 - 1	148	
6 - 1	148.6	
7 - 1	148.7	
10 - 1	153.9	
11 - 1	145	
12 - 1	146.9	
13 - 1	151.4	
14 - 1	152.3	
15 - 1	147.3	
17 - 1	149	
18 - 1	145.2	

Westgard	تفسیر	Transferrin	آزمایش	BS-480 اتوآنالایزر	دستگاه	4	ماه	1402	سال
Control	biorex	Mean	148.96	SD	3.14	Exp. Date	1402/5/1	Lot. No	641-01-030/2
Next Month	Mean	148.86	SD	3.26	CV	2.19			

