

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.65	
Mean	172.44	Mean	227.06
SD	2.33	SD	3.75
CV	1.35	CV	1.65
Day-Run	Value	Day-Run	Value
1 - 1	175	1 - 1	222
3 - 1	174	3 - 1	230
4 - 1	170	4 - 1	226
5 - 1	174	5 - 1	231
6 - 1	173	6 - 1	224
10 - 1	170	10 - 1	226
11 - 1	171	11 - 1	229
13 - 1	173	13 - 1	227
19 - 1	171	19 - 1	230
20 - 1	176	20 - 1	230
21 - 1	175	21 - 1	223
24 - 1	175	24 - 1	220
25 - 1	172	25 - 1	230
26 - 1	175	26 - 1	235
27 - 1	168	27 - 1	226
28 - 1	170	28 - 1	223
29 - 1	170	29 - 1	228
31 - 1	172	31 - 1	227

Westgard	تفسیر	Triglycerides	آزمایش	اتوآنالایزر BS-800	دستگاه	4	ماه	1402	سال
Normal Control	biorex	Mean	172.1	SD	2.47	Exp. Date	1403/1/17	Lot. No	035

