

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 2.27	
Mean	4.3	Mean	9.02
SD	0.07	SD	0.2
CV	1.6	CV	2.27
Day-Run	Value	Day-Run	Value
1 - 1	4.3	1 - 1	8.7
3 - 1	4.3	3 - 1	9.2
4 - 1	4.2	4 - 1	8.8
5 - 1	4.3	5 - 1	9.1
6 - 1	4.3	6 - 1	9.1
10 - 1	4.3	10 - 1	9.14
11 - 1	4.3	11 - 1	9
13 - 1	4.3	13 - 1	9.1
19 - 1	4.3	19 - 1	9.2
20 - 1	4.4	20 - 1	9.1
21 - 1	4.4	21 - 1	8.8
24 - 1	4.3	24 - 1	8.7
25 - 1	4.4	25 - 1	9
26 - 1	4.4	26 - 1	9.5
27 - 1	4.2	27 - 1	8.9
28 - 1	4.2	28 - 1	8.8
29 - 1	4.2	29 - 1	9.1
31 - 1	4.3	31 - 1	9.1

Westgard	تفسیر	Uric Acid	آزمایش	اتوانالایزر BS-800	دستگاه	4	ماه	1402	سال
Normal Control	biorex	Mean	4.26	SD	0.15	Exp. Date	1403/1/17	Lot. No	035

