

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.06	
Mean	34.58	Mean	133.61
SD	0.93	SD	2.75
CV	2.69	CV	2.06
Day-Run	Value	Day-Run	Value
1 - 1	34	1 - 1	138.5
2 - 1	35	2 - 1	133.8
3 - 1	34.5	3 - 1	135
6 - 1	34	6 - 1	130
8 - 1	35.7	8 - 1	132.5
9 - 1	36	9 - 1	132
10 - 1	36.5	10 - 1	133.8
11 - 1	34	11 - 1	129.9
17 - 1	34.1	17 - 1	131
20 - 1	34.7	20 - 1	133
21 - 1	35	21 - 1	128.6
22 - 1	33	22 - 1	134.5
23 - 1	35.1	23 - 1	138.5
24 - 1	34.8	24 - 1	134.5
25 - 1	33.5	25 - 1	134.2
27 - 1	35.3	27 - 1	136.5
28 - 1	33	28 - 1	133.1
29 - 1	34.8	29 - 1	131.4
30 - 1	33.7	30 - 1	133.9
31 - 1	34.8	31 - 1	137.5

Westgard	تفسیر	Urea	آزمایش	BS-800 اتوآنالایزر	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	34.5	SD	1.27	Exp. Date	1403/1/17	Lot. No	035

